

SURGE CURRENT GENERATOR

PG 6 - 400

PG 6 - 200

Surge current

8 / 20 µs

0.4 - 10 kA 0.2 - 5 kA

acc. to IEC 60060, VDE 0432



The high current pulse generators PG 6-400 / PG 6-200 are designed for surge testing of electrical components, over-voltage protectors and electronic circuits acc. to IEC, VDE etc. They generate standard impulse currents with waveform $8/20~\mu s$. Short circuit output current amplitude is selectable from 0.2 up to 10 kA by adjustment of the charging voltage of the internal energy storage capacitor.

The polarity of the output current is selectable. Positive, negative or alternating polarity of the output current can be pre selected.

The pulse-forming network contains a high pulse-fidelity current viewing resistor for measurement of the output current amplitude and waveform with a scope.

The impulse current output is located at the top of the equipment and provides high-current connectors for a plug-in test adapter. A dielectric cover with safety interlock protects the high-voltage output terminals. Upon lifting of the cover, switching-off of the generator or mains blackout a built-in high-voltage grounding switch discharges the test object and the internal energy storage capacitor.

The generator excels by its compact design, simple handling and precise reproducibility of test impulses. It features a microprocessor controlled user interface and a 5" touch screen unit for ease of use. The microprocessor allows the user to execute either standard test routines or a "user defined" test sequence. A standard USB port provides the ability to print a summary of the test parameters to a USB stick.

The software program IPG-REMOTE allows full remote control of the test generator via Ethernet light guide as well as documentation and evaluation of test results, accordingly to the IEC 17025. To record definite impulses, it is equipped with an Impulse Recording Function (IRF). Moreover all generator functions may be computer controlled via the isolated optical interface.



| TECHNICAL SPECIFICATIONS | PG 6-200 / PG 6-400 |
|--|------------------------------------|
| Mainframe | |
| Microprocessor controlled touch panel | 5", 800X480, 24 bit |
| Optical Ethernet Interface for remote control of the generator | optional |
| Interface for saving reports | USB |
| Connector for external safety interlock loop | 24 V = |
| External trigger input /output | 10 V at 1 kΩ |
| External red and green warning lamps | 230 V, 60W |
| Mains power | 230 V, 50/60 Hz |
| Dimensions of desk top case W * H * D | 450*185*500 mm ³ |
| Weight | 25 kg |
| Pulse forming network: surge current 8/20 µs | 5 kA / 10 kA |
| Charging voltage, adjustable | 0.2 - 6.25 kV |
| | |
| Surge current, (short circuit at the output) PG 6-400 | 0.4 - 10 kA ± 10 % |
| Surge current, (short circuit at the output) PG 6-200 | 0.2 - 5 kA ± 10 % |
| | |
| Waveform acc. to IEC 60060, VDE 0432 | 8 / 20 µs ± 20% |
| Polarity of output current, selectable | pos / neg /alt. |
| | |
| Max. stored energy | 200 / 400 Joule |
| Charging time for max. charging voltage approx. | 10s / 20 s |
| | |
| High current output terminals | 6 mm Ø, connector |
| | |
| Monitor output for output current PG 6-400 | $10V = 10 \text{ kA } \pm 5 \%$ |
| Monitor output for output current PG 6-200 | $10V \equiv 5 \text{ kA} \pm 5 \%$ |
| | |
| Safety test cover | |
| mounted on the top of the equipment, | |
| safety interlock loop connected to the limit switch, | |
| red and green warning lamps installed | |
| Dimensions W * H * D | 400 * 150 * 250 mm ³ |
| Ontion | |
| Option DC Software for remote control of the generator (VD 7.8) | |
| PC Software for remote control of the generator, (XP,7,8). | |
| PC Ethernet Interface USB/RS232, optically isolated and light | |
| guide, 5 m long. | |
| Monitor output for output voltage | ratio 1000:1 ± 20/ |
| Monitor output for output voltage | ratio 1000:1 ± 2% |