

# PEARSON™ CURRENT TRANSFORMER MODEL 5550

Nominal current division ratio	1:100
Maximum rms current	500 Amperes
Nominal external terminating resistance	0.016 Ω
Ratio error at 20 kHz	±1% maximum
Phase shift at 20 kHz	±1° maximum
Low frequency limit at 500 Amperes rms	10 Hz
Ratio error at low frequency limit	±1% maximum
Phase shift at low frequency limit	±1° maximum
Output terminals	#8-32 screw terminals

This model is designed for use with the Yokogawa 2533 Digital Power Meter, and should be used for currents up to 500 Amperes, on the 5 Ampere meter setting.

