


---

# FREQUENCY RESPONSE PHASE SHIFT

---



Pearson Current Monitors exhibit amplitude response and phase shift as a function of frequency similar to that depicted in the above figure. To apply this graph to a specific model, use the low and high 3 dB points given in the specification sheet.